

3-A CERTIFIED SANITARY THERMOWELLS

#1	SERIES										
5F	Sanitary Thermowells - Add "W" here for a plug with a chain attached to well. (i.e. 5FW)										
#2	STYLE [25-27]										
A	Step shank	F	Fast response straight shank (1/2" Q)				S	Straight shank (Standard)		T	Tapered shank
#3	BORE SIZE										
2	.260" ID				3	.385" ID			X	Other, specify	
#4	U (INSERTION) DEPTH [15]					"U" DIMENSION		SENSOR LENGTH			
B	2 1/2"					2 1/2"		4"			
C	4 1/2"					4 1/2"		6"			
D	6"					6"		7 1/2"			
E	7 1/2"					7 1/2"		9"			
X	Other, specify										
#5	CAP SIZE [SECTION 4]										
05	1/2" x 3/4"	20	2"	30	3"	60	6"	100	10"	X	Other, specify*
15	1" x 1 1/2"	25	2 1/2"	40	4"	80	8"	120	12"		
#6	BRAND OF CAP										
T	Tri-Clover (16 AMP)				I	I Clamp (16Al-14l)**				A	3A4
B	Bevel Seat w/o 13H Nut***				P	PV Gasket (16APV) w/o 13-H Nut				X	Adapter***
BH	Bevel Seat w/ 13H Nut***				PH	PV Gasket (16APV) w/ 13-H Nut				X	Other,specify*
#7	MATERIAL										
H	304 SS							L	316L SS		
I	304L SS							X	Other, specify*		
K	316 SS										
#8	POLISH										
H	High polish (#4 Finish ≤ 32 Microinches) Standard										
E	Electropolish or (8G Finish ≤ 8 Microinches)										
P	Passivate (After #4 Finish)										
#9	TAGGING OPTIONS										
1	Stamped on Well (Standard)										
X	Other Note: You must always specify information required on tag										
Z	N/A										
#10	DOCUMENTATION / CERTIFICATION - -Use all that apply (I.E. "DU" requests Dye Penetrant Test and X-Ray Examination)										
M	Material Test Reports (MTR)										
D	Dye Penetrant Testing										
P	Internal Hydrostatic Pressure Test										
U	X-Ray Examination										
W	PTC 19.3 Strength Calculation (Murdock Calculation)										
S	Surface Finish Certificate										
E	Certificate of Electropolish										
A	Certificate of No Animal Derived Material (ADM)										
N	Certificate of No Polishing Compounds										
O	Certificate of Cleaned for Oxygen Service										

5F	A	2	D	20	T	K	H	1	M
----	---	---	---	----	---	---	---	---	---

*When specifying "X" ensure that it meets 3A standard.
 **Not 3A authorized.
 ***Must be cleaned manually.